



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**North Texas Scales, Inc.
2727 N. Highway 175
Seagoville, TX 75159**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

**Calibration of Industrial Weighing Equipment
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Draft

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

<i>Initial Accreditation Date:</i>	<i>Issue Date:</i>	<i>Revision Date:</i>	<i>Expiration Date:</i>
June 18, 2008	May 13, 2010	July 20, 2011	May 12, 2012

<i>Accreditation No:</i>	<i>Certificate No:</i>	<i>Page No:</i>
62602	L10-73-R1	Page 1 of 2



Certificate of Accreditation: Supplement

North Texas Scales, Inc.
2727 N. Highway 175
Seagoville, TX 75159

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Class II Scales	10 mg to 10 000 g	$(1.15 \times 10^{-2} + 1.15 \times 10^{-4}Wt)$ g	Class F Weights NIST Handbook 44
Class III Scales	5 lb to 20 000 lb	$(2.31 + 4.8 \times 10^{-5}Wt)$ lb	
	20 001 lb to 50 000 lb	$(4.92 + 6.5 \times 10^{-5}Wt)$ lb	
Class IIIIL Scales	10 lb to 100 000 lb	$(11.55 + 4.8 \times 10^{-5}Wt)$ lb	
	100 001 lb to 200 000 lb	$(18.92 + 6.9 \times 10^{-5}Wt)$ lb	

1. This laboratory offers commercial calibration service and on-site calibration service.
2. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
3. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.